

=====

Acq. Operator	: SYSTEM	Seq. Line	: 1
Acq. Instrument	: LCMS-1	Location	: P1-C9
Injection Date	: 9/15/2020 10:09:47 AM	Inj	: 1
		Inj Volume	: 2.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.100 µl

Acq. Method : C:\CHEM32\1\METHODS\PFAAs.M
Last changed : 5/6/2019 8:49:48 AM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 9/14/2020 4:07:19 PM by SYSTEM
(modified after loading)

Method Info : Operator: Oraganic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Current Chromatogram(s)

