



=====

Acq. Operator : SYSTEM	Seq. Line : 22
Sample Operator : SYSTEM	
Acq. Instrument : LCMS-2	Location : P1-B-08
Injection Date : 10/26/2022 6:23:13 PM	Inj : 2
	Inj Volume : 2.000 µl
Different Inj Volume from Sample Entry!	Actual Inj Volume : 0.500 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\10262201\10262201 2022-10-26 11-45-48\PFAM.M

Last changed : 1/26/2021 10:30:23 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M

Last changed : 10/24/2022 10:17:59 AM by SYSTEM
(modified after loading)

Method Info : OPERATOR: Organic QC lab
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

