



Acq. Operator : SYSTEM Seq. Line : 3

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2 Location : P2-A6

Injection Date : 10/22/2020 3:15:28 PM Inj : 1

Inj Volume : 2.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.100 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\10222001\10152001 2020-10-22 14-36-52
\PFAA.M

Last changed : 4/27/2020 2:54:36 PM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 10/20/2020 2:10:33 PM by SYSTEM
(modified after loading)

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Current Chromatogram(s)

