



=====
Acq. Operator : SYSTEM Seq. Line : 45

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2 Location : P1-A1

Injection Date : 6/7/2022 8:32:26 PM Inj : 3

Inj Volume : 2.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\06072201\06072201 2022-06-07 06-39-34
\PFAA.M

Last changed : 1/26/2021 10:30:23 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M

Last changed : 5/16/2022 12:35:31 PM by SYSTEM
(modified after loading)

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min,the 100%
Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system
, Agilent 6125 quadrupole MS detector.

Current Chromatogram(s)

