



```

=====
Acq. Operator   : SYSTEM                      Seq. Line :    3
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                      Location  : P1-B1
Injection Date  : 6/23/2021 7:04:06 AM        Inj       :    1
                                                Inj Volume: 2.000 µl
    
```

Different Inj Volume from Sample Entry! Actual Inj Volume : 1.000 µl

```

Acq. Method     : C:\Users\Public\Documents\ChemStation\1\Data\06232101\06222101 2021-06-23 06-25-22
                  \PFAA.M
    
```

Last changed : 1/26/2021 10:30:23 AM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

