



=====

Acq. Operator : SYSTEM	Seq. Line : 2
Sample Operator : SYSTEM	
Acq. Instrument : LCMS-2	Location : P1-D-07
Injection Date : 7/16/2021 6:15:12 AM	Inj : 2
	Inj Volume : 2.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.100 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\07162101\07152101 2021-07-16  
05-55-20\PFAM.M  
Last changed : 1/26/2021 10:30:23 AM by SYSTEM  
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity\_195-315.M  
Last changed : 4/29/2021 10:41:07 AM by SYSTEM  
Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

