



=====

| | |
|--------------------------------------|-----------------------|
| Acq. Operator : SYSTEM | Seq. Line : 2 |
| Sample Operator : SYSTEM | |
| Acq. Instrument : LCMS-2 | Location : P2-F-03 |
| Injection Date : 1/9/2023 7:36:30 AM | Inj : 2 |
| | Inj Volume : 2.000 µl |

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.100 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\01092301\010923101 2023-01-09 07-16-43\PF00.M

Last changed : 1/26/2021 10:30:23 AM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 3/23/2022 11:04:54 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Current Chromatogram(s)

