



```

=====
Acq. Operator   : SYSTEM                               Seq. Line :    3
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                               Location  :   P2-A-01
Injection Date  : 6/22/2022 6:49:01 AM                Inj       :    1
                                                    Inj Volume: 2.000 µl

```

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.010 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\06222201\06222201 2022-06-22 06-10-25\PFAS.M

Last changed : 1/26/2021 10:30:23 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M

Last changed : 3/23/2022 11:04:54 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

