



Acq. Operator : SYSTEM Seq. Line : 25

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2 Location : P1-E8

Injection Date : 5/16/2023 3:46:08 PM Inj : 1

Inj Volume : 2.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.200 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\051623\051623101 2023-05-16 08-11-32\PFSA.M

Last changed : 1/26/2021 10:30:23 AM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Current Chromatogram(s)

