


File :D:\MassHunter\GCMS\1\DATA\072420\09- -KD.D  
Operator : Organic QC Lab  
Acquired : 24 Jul 2020 08:12 using AcqMethod CHICK\_2019\_2S\_2uL.M  
Instrument : GCMS 6  
Sample Name: PFRS-018S  
Misc Info :  
Vial Number: 146

 **AccuStandard**<sup>®</sup>  
Leader in Analytical Reference Standards  
Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

