


File : L:\QC Data\GCMS-06 Minimal\DATA\072221\12- ZA.D
Operator : Organic QC Lab
Acquired : 22 Jul 2021 15:58 using AcqMethod CHICK_2019.M
Instrument : GCMS 6
Sample Name: PHTH-D4-011S
Misc Info :
Vial Number: 112



AccuStandard®
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

