

File :Q:\GCMS-08 Minimal\DATA\20230324\24- [ JP ].D  
Operator : Organic QC Lab  
Acquired : 24 Mar 23 08:59 pm using AcqMethod Hy\_BDE\_24cm\_S1.M  
Instrument : GCMS 8  
Sample Name: PLAS-AX-041S  
Misc Info : Cyanox LTDP; in hexane  
Vial Number: 13



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

