

File :D:\MassHunter\GCMS\1\data\20230628\07 -JP-.D
Operator : Organic QC Lab
Acquired : 28 Jun 23 03:15 pm using AcqMethod Hy_BDE_24cm_S1.M
Instrument : GCMS 8
Sample Name: PLAS-AX-069S
Misc Info : Irganox 1035; in hexane
Vial Number: 5



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

