

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-08 Minimal\DATA\20221013  
... \34- KM.D  
Operator : Organic QC Lab  
Instrument : GCMS 8  
Acquired : 13 Oct 2022 23:32 using AcqMethod Hy\_Pest\_15cm\_S10.M  
Sample Name: PLAS-AX-0855  
Misc Info : Ethanox 4703 in Hexane



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

