

File : Q:\GCMS-08 Minimal\DATA\20230814\80- (JP).D  
Operator : Organic QC Lab  
Acquired : 14 Aug 2023 15:03 using AcqMethod Hy\_PLASCAL-001\_S2\_mz40-1050.M  
Instrument : GCMS 8  
Sample Name: PLAS-CAL-001  
Misc Info : in IPrOH  
Vial Number: 6



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

