

File : Q:\GCMS-08 Minimal\DATA\20200612\06-DC.D
Operator : Organic QC Lab
Acquired : 12 Jun 20 01:11 pm using AcqMethod Hy_Pest_15cm_S50.M
Instrument : GCMS 8
Sample Name: PLAS-PL-015S
Misc Info : Benzoflex(R) 2-45 in Hexane
Vial Number: 4



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 24c/min to 340c, 2min
GC Parameters: Cons. Split, 5.4psi constant flow
Split 50:1 1uL inj.; GC/MS; INJ 270c

