

File :Q:\GCMS-08 Minimal\DATA\20230627\25- [JP].D
Operator : Organic QC Lab
Acquired : 27 Jun 23 07:02 pm using AcqMethod Hy_Pest_24cm_S10.M
Instrument : GCMS 8
Sample Name: PLAS-PL-075S-A
Misc Info : Diisodecyl azelate; in acetone
Vial Number: 9



Column: DB-5, 30m, 0.25ID, 0.25um
Oven: 80c, 2m, 30c/m to 320, final 2min
GC para: Pulse split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj, GCMS, Inj 280c

