

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-08 Minimal\DATA\20230615
... A\06- KM.D
Operator : Organic QC Lab
Instrument : GCMS 8
Acquired : 15 Jun 2023 16:47 using AcqMethod Hy_PAH_24cm_S50.M
Sample Name: PLAS-PL-097S
Misc Info : Methyl oleate in Hexane



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

