

Acq. Operator : SYSTEM

Seq. Line :

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2

Location :

Injection Date : 1/31/2025 1:34:30 PM

Inj :  $\perp$

Inj Volume : 1.000  $\mu$ l



**AccuStandard**  
Leader in Analytical Reference Standards

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\013125\013125 2025-01-31 06-38-52  
\Purity\_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity\_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8  $\mu$  particlue size. Agilent 1290 Infinity II LC/MS system , Agilent 6125 quadrupole MS detector.

Sample Info : Dipropylene glycol dibenzoate in Hexane

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

