

```
=====
Acq. Operator   : SYSTEM                               Seq. Line :    5
Acq. Instrument : LCMS-1                             Location  : P1-E1
Injection Date  : 3/14/2023 4:45:52 PM              Inj       :    1
                                                    Inj Volume: 1.000 µl
```



```
Acq. Method    : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 1/5/2023 12:01:53 PM by SYSTEM
Analysis Method: C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 3/8/2023 10:15:58 AM by SYSTEM
                (modified after loading)
Method Info    : Operator: Organic QC lab
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size.
                Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV
                Reference.
```

Sample Info : 2-(dodecylthiocarbonothioylthio)-2-methyl propionic acid @ 1000 ug/mL in Hexane

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

