

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-06 Minimal\DATA\031822\3
... 3- KM.D
Operator : Organic QC Lab
Instrument : GCMS 6
Acquired : 18 Mar 2022 19:40 using AcqMethod CHICK_2019.M
Sample Name: PLAS-ST-002S
Misc Info : RT: Stearic acid in Hexane



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

