

File : D:\MassHunter\GCMS\1\data\06292023\02- ALR.D
Operator :
Acquired : 29 Jun 2023 10:55 using AcqMethod Hy SV DB-5MS UI20M 0.18 40 split Time 14 MIN .M
Instrument : GCMS 9
Sample Name: PLAS-UV-009S-CN
Misc Info :
Vial Number: 1



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

