

File :L:\QC Data\GCMS-08 Minimal\DATA\20220509\07 - SAK.D
Operator : Organic QC Lab
Acquired : 09 May 2022 13:21 using AcqMethod Hy_PAH_24cm_S100.M
Instrument : GCMS 8
Sample Name: PLAS-UV-012S-CN
Misc Info : resample
Vial Number: 125



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

