

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-06 Minimal\DATA\021122\7  
... 1- KM.D  
Operator : Organic QC Lab  
Instrument : GCMS 6  
Acquired : 12 Feb 2022 09:01 using AcqMethod CHICK\_2019\_13MIN.M  
Sample Name: PLAS-UV-013S-CN  
Misc Info : 2-(2-Hydroxy-5-tert-octylphenyl)benzotriazole @ 1000 ug/mL in AcCN



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

