

File :Q:\GCMS-06 Minimal\DATA\082619\BSB\25-.D
Operator : D.C
Acquired : 27 Aug 2019 00:48 using AcqMethod CHICK_2019.M
Instrument : GCMS 6
Sample Name: PS-590D-06
Misc Info : Aroclor 1254 in MeOH
Vial Number: 52



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

