

File :D:\MassHunter\GCMS\1\data\03292023\25-ALR.D  
Operator :  
Acquired : 29 Mar 2023 16:48 using AcqMethod ALR 19 min 20 split FAST .M  
Instrument : GCMS 9  
Sample Name: R-050S  
Misc Info : 2,6-Dinitrotoluene in Toluene  
Vial Number: 10



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm  
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

