

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-07 Minimal\DATA\04212301  
... \36- KM.D  
Operator : Organic QC department  
Instrument : 5977B MSD  
Acquired : 22 Apr 2023 01:04 using AcqMethod KERLA-2017\_Short\_15Min.M  
Sample Name: R-087S  
Misc Info : 2,4-Dichloronitrobenzene in Toluene



**AccuStandard®**

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

