

File : Q:\GCMS-08\DATA\20181120\39-  
Operator : JJP / KM  
Acquired : 21 Nov 18 05:19 pm using AcqMethod KERLA-2017-100uL draw speed.M  
Instrument : GCMS 8  
Sample Name: S-10469  
Misc Info : Methyl mercury chloride in MeOH  
Vial Number: 14



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

