

```
=====
Acq. Operator   : SYSTEM                      Seq. Line :    3
Acq. Instrument : LCMS-1                      Location  : P2-B2
Injection Date  : 8/2/2019 7:27:37 AM        Inj       :    1
                                                Inj Volume: 1.000 µl
```

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.500 µl

```
Acq. Method      : D:\DATA\0802901\08021901 2019-08-02 06-47-29\Purity195_315.M
Last changed     : 8/2/2019 6:47:29 AM by SYSTEM
Analysis Method  : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed     : 8/1/2019 9:30:32 AM by SYSTEM
                  (modified after loading)
```

```
Method Info      : Operator: GA
                  Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                  Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                  :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                  particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
                  MS Detector. No UV Reference.
```

Sample Info : Metazosulfuron

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

