

File : L:\QC Data\GCMS-06 Minimal\DATA\062323\53 - sak.D  
Operator : Organic QC Lab  
Acquired : 24 Jun 2023 08:19 using AcqMethod CHICK\_2019\_VOC-SV\_S100.M  
Instrument : GCMS 6  
Sample Name : SWMO-70X-100ML  
Misc Info : diluted  
Vial Number : 46



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 100:1, 1µl inj.; GC/MS; Injector temp 270°C

