

File : L:\QC Data\GCMS-08 Minimal\DATA\20230602\25 - sak.D  
Operator : Organic QC Lab  
Acquired : 02 Jun 2023 20:52 using AcqMethod Hy\_VOC-SV\_24cm\_S50.M  
Instrument : GCMS 8  
Sample Name: SSWMO-LT-50X-100ML  
Misc Info : dil 1:10 with Isooctane  
Vial Number: 112



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 35°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

