

File : L:\QC Data\GCMS-08 Minimal\DATA\20230602\25 - sak.D
Operator : Organic QC Lab
Acquired : 02 Jun 2023 20:52 using AcqMethod Hy_VOC-SV_24cm_S50.M
Instrument : GCMS 8
Sample Name: SWMO-LT-5X-100ML
Misc Info : dil 1:10 with Isooctane
Vial Number: 112



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 35°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

