

File : Q:\GCMS-08 Minimal\DATA\20190423\23- D
Operator : JP/KM
Acquired : 24 Apr 2019 00:47 using AcqMethod KERLA-VOC-SV.M
Instrument : GCMS 8
Sample Name: UCMR4-M-552.3 Derivatized
Misc Info : HAA Standard in MtBE Derivatized
Vial Number: 133



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

