

File : L:\QC Data\GCMS-09 Minimal\DATA\01122023\25 - Pak.D  
Operator :  
Acquired : 12 Jan 2023 19:45 using AcqMethod ALR 19 min 100 split FAST .M  
Instrument : GCMS 9  
Sample Name: UFA-029S  
Misc Info :  
Vial Number: 104



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm  
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

