

File :Q:\GCMS-06 Minimal\DATA\020921\19- KM.D
Operator : Organic QC Lab
Acquired : 09 Feb 2021 19:13 using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name:UOP-M-660
Misc Info : UOP Method 660 Standard in Water
Vial Number: 108



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

