

File : Q:\GCMS-06 Minimal\DATA\062819\15-.D  
Operator : D.C  
Acquired : 28 Jun 2019 20:46 using AcqMethod CHICK\_2019.M  
Instrument : GCMS 6  
Sample Name: USP-561-04  
Misc Info : USP 561 Pest Std in AcCN  
Vial Number: 59



Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

