

File : L:\QC Data\GCMS-06 Minimal\DATA\051222\10 - SAK.D  
Operator : Organic QC Lab  
Acquired : 12 May 2022 14:47 using AcqMethod CHICK\_2017\_13MIN\_S100.M  
Instrument : GCMS 6  
Sample Name : V-028S-D-10X  
Misc Info :  
Vial Number : 114



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

