

File : L:\QC Data\GCMS-08 Minimal\DATA\20220603\02 - SAK.D
Operator : Organic QC Lab
Acquired : 03 Jun 2022 08:07 using AcqMethod Hy_PAH_24cm_S100.M
Instrument : GCMS 8
Sample Name: Z-014J-3-M-0.5X
Misc Info :
Vial Number: 34



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

