



AccuStandard®

File :M:\GCMS-07 Minimal\DATA\06092101\31-ALR.D
Operator : Organic QC department
Acquired : 09 Jun 2021 20:14 using AcqMethod KERLA-2017_Short_15Min_S100.M
Instrument : 5977B MSD
Sample Name: Z-014J-4
Misc Info :
Vial Number: 74

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

