

File :Q:\GCMS-04 Minimal\DATA\051519\BSB\09-.D
Operator : KM
Acquired : 15 May 2019 19:42 using AcqMethod VOC-100SPLIT DECKER.M
Instrument : GCMS4
Sample Name: AE-00036-CAL-01
Misc Info :
Vial Number: 27



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

