


File : L:\QC Data\GCMS-06 Minimal\DATA\100220\18- ZA.D
Operator : Organic QC Lab
Acquired : 02 Oct 2020 19:43 using AcqMethod CHICK_2019_VOC-SV.M
Instrument : GCMS 6
Sample Name: ALR-008S-ET-10X
Misc Info :
Vial Number: 136

 **AccuStandard**[®]
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

