



File :K:\QC Data\GCMS-04 Minimal\DATA\040820\13-.D
Operator : RFXM
Acquired : 8 Apr 2020 17:23 using AcqMethod VOC-100SPLIT DECKER.M
Instrument : GCMS4
Sample Name: ALR-170S
Misc Info :
Vial Number: 30

Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

