

File : Q:\GCMS-06 Minimal\DATA\061821\18- -DC.D
Operator : Organic QC Lab
Acquired : 18 Jun 21 07:00 pm using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name: APP-9-016-10X
Misc Info : Benz(a)anthracene in Methanol
Vial Number: 141

AccuStandard
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

