



AccuStandard®

File : M:\GCMS-06 Minimal\DATA\052821\10-ALR.D
Operator : Organic QC Lab
Acquired : 28 May 2021 14:02 using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name: APP-9-017
Misc Info : Benzo(b)fluoranthene in Methanol
Vial Number: 36

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

