

File :M:\GCMS-06 Minimal\DATA\061021\34- ALR.D  
Operator : Organic QC Lab  
Acquired : 11 Jun 2021 00:10 using AcqMethod CHICK\_2019\_S100.M  
Instrument : GCMS 6  
Sample Name: APP-9-020-D-20X  
Misc Info : Benz[a]pyrene in MeCl2  
Vial Number: 83



**AccuStandard®**

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

