

File :D:\MassHunter\GCMS\1\DATA\01192101\12- -KD.D
Operator : Organic QC department
Acquired : 19 Jan 2021 17:32 using AcqMethod KERLA-2017.M
Instrument : 5977B MSD
Sample Name: APP-9-047
Misc Info :
Vial Number: 63



Column: DB-5MS, 30m, 0.25 ID, 0.25 μ m
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1 μ L inj.; GC/MS; INJ 270c

