

File : Q:\GCMS-08 Minimal\DATA\20230714\50- JP=.D  
Operator : Organic QC Lab  
Acquired : 15 Jul 23 12:01 pm using AcqMethod Hy\_PAH\_24cm\_S150.M  
Instrument : GCMS 8  
Sample Name: APP-9-076-M-50X  
Misc Info : 2,6-Dichlorophenol; in meoh  
Vial Number: 41



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

