



AccuStandard®

File :D:\MassHunter\GCMS\1\DATA\04232101B\25- ALR.D
Operator : Organic QC department
Acquired : 23 Apr 2021 15:52 using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name APP-9-153
Misc Info : N-Nitrosomorpholinein MeCl2
Vial Number: 36

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

