

File : Q:\GCMS-04 Minimal\DATA\060619\01-.D
Operator : RFXM/KM
Acquired : 6 Jun 2019 15:10 using AcqMethod VOC-100SPLIT DECKER.M
Instrument : GCMS4
Sample Name: AS-E0002
Misc Info :
Vial Number: 42



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

