

File :Q:\GCMS-08 Minimal\DATA\20191113\BSB\05-.D  
Operator : KM / JP / ZA  
Acquired : 13 Nov 2019 13:58 using AcqMethod KERLA-2017\_S100.M  
Instrument : GCMS 8  
Sample Name: AS-E0026  
Misc Info : 3,3'-Dichlorobenzidine in Methanol  
Vial Number: 31



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

